

INTRODUCTION

Manual radiation therapy plan checking is essential for ensuring treatment accuracy and patient safety but can be time-consuming and subject to reviewer variability.

To improve consistency and efficiency, we developed a deterministic, RayStation-integrated automated plan-check system that performs more than 30 checks across multiple clinical domains using predefined acceptance criteria and transparent reporting of measured values. The system aims to reduce reviewer variability, streamline workflow, and enhance patient safety through standardized plan-review automation.

METHODS AND MATERIALS

An automated plan-check system was developed using the RayStation (2024A) Python API. The framework utilizes a modular QA engine in which all checks are implemented using deterministic pass/fail logic with predefined acceptance thresholds. Each check evaluates TPS-accessible parameters and reports both the measured value and the corresponding acceptance criterion within a standardized HTML report. The system performs more than 30 automated checks across eight major clinical domains:

- CT & Imaging
- Prescription
- Machine & Dose Grid
- Approval
- Structure Validation
- Beam & Geometry
- Plan Quality
- Isocenter & Setup

Implemented checks include verification of simulation parameters, slice thickness, CT dataset selection, prescription consistency, beam model assignment, dose-grid resolution, structure-type validation, beam geometry consistency, arc-direction naming, modulation limits, target coverage, hotspot evaluation, clearance checks, setup-beam consistency, and unintended isocenter deviations.

Yinghua Tao, Ph.D, DABR
University of Washington
Email: ytao7@uw.edu
Phone: 206-606-6724

The automated plan-check system successfully processed over 200 clinical treatment plans, including VMAT, SBRT, and 3D conformal plans. Average runtime was under 3 seconds per plan, enabling near real-time plan-review support within the clinical workflow.

The system reliably identified predefined failing conditions across multiple QA domains, including:

CT Imaging		Pass	Fail
Status	Message	Actual	Expected
Pass	CT date check passed	12012025	Matches user input: 12012025
Pass	CT slice number check passed	142	Matches user input: 142
Pass	CT read date operation passed	10/5 vs 10/4/10/20/26	10/5 vs 10/4/10/20/26
Pass	Localization check passed	(0.15, -6.85, 0.00m)	Localization file exists and has valid coordinates
Precision		Pass	Fail
Status	Message	Actual	Expected
Pass	Beamsort V Brd AP/RT facelock check passed	5	Matches user input: 5
Pass	Beamsort V Brd AP/RT facelock done check passed	6001 c/sy	Matches user input: 600 c/sy
Pass	Beamsort V Brd AP/RT acc = V PTY check passed	Structure: PTY_3000 (Prescribed: 3000 c/sy @ 95.0% Activated: 3011 c/sy)	Expected: PTY structure with Dose/Volume > 95%
Machine & Dose grid		Pass	Fail
Status	Message	Actual	Expected
Pass	Dose grid loaded	5 - Scales commissioned at 7/25/2021 13:22 PM	Expected: Scales commissioning time: 7/25/2021 13:22 PM
Pass	Dose grid resolution check passed	(0, 0.2, 0.2mm)	< 0.2 mm in all directions
Pass	All structures are within the dose grid	Structure bounding box: min=[27.6, -27.6, -33.3]m; max=[26.5, 17.2, 42.8]m	Dose grid bounds: min=[27.6, -27.6, -33.3]m; max=[28.4, 17.5, 44.4]m
Approval		Pass	Fail
Status	Message	Actual	Expected
Pass	Beamsort V Brd AP/RT approval check passed	Beamsort approved	Beamsort must be approved
Pass	Plan approval check passed	Review = AMCB465, Time = 12/22/2025 12:36:25 PM	Plan must be approved
Structure checks		Pass	Fail
Status	Message	Actual	Expected
Pass	Target type check passed	PTV (user selected) must be of type 'Target'	External SCL with valid contours exists
Pass	External contour check passed	EXTERNAL	External SCL with valid contours exists
Fail	No valid couch support structures found in structure set	Target 0 valid couch support RCLs. Present Support RCLs: none	Target 0 valid couch support RCLs with defined material: 20x20x20 density, and valid contours. Valid names: BEAM-00 Couch Couch, BEAM-00 Couch Couch, Target Target Couch Exterior, Varian iCST Couch Interior, ckt-MVA Couch
Beam & Geometry		Pass	Fail
Status	Message	Actual	Expected
Pass	Collimator angle check passed for Beam 1 C550xG105 R Brd in beamsort V Brd AP/RT	Collimator angle = 7.30°	Should not be 0° for arcs
Pass	Collimator angle check passed for Beam 2 C105xG50 R Brd in beamsort V Brd AP/RT	Collimator angle = -33.00°	Should not be 0° for arcs
Pass	Beam 1 description angles match plan (1 PT)	Name: C550xG105 R Brd, start=55.0°, step=105.0°	Start and stop angles match between beam description and actual angles
Pass	Beam 2 description angles match plan (1 PT)	Name: C105xG50 R Brd, start=105.0°, step=55.0°	Start and stop angles match between beam description and actual angles
Pass	Beam 1 COW ramping matches arc in V Brd AP/RT	Test: V550xR105 R Brd, Acc=COW	Test direction matches Arc/Rotation/Direction
Pass	Beam 2 COW ramping matches arc in V Brd AP/RT	Test: V105xG50 R Brd, Acc=COW	Test direction matches Arc/Rotation/Direction
Pass	Couch tick check passed	Angles: 0.0°, 0.0°	No small unrounded ticks (< 0°)
Plan Quality		Pass	Fail
Status	Message	Actual	Expected
Pass	Beamsort 'Skullbase' modulation check passed	1.43	Modulation factor < 5
Pass	Normalization check passed (non-SBRT)	1.06	0.95 < MaxMU < 1.15
Fail	Max dose not in PTV	PTV Max: 7181.1 External Max: 3253.2	PTV MaxMU must be equal to External Max (2° plan can be an exception)
Fail	Target coverage insufficient	V100 = 69.6%, V90 = 100.0%	V100 > 95%, V90 > 98%
Fail	Grid not used for RCL PTY in beamsort 'Skullbase'	Estimated 201.96, Target Dose/Volume @ 0.0% = -68.6m 3000.00	All defined contour gains must be met

Figure 1. Example Plan Check Report. Demonstrates consistent, reviewer-independent error detection across **eight key domains**: CT & imaging, prescription, machine & dose grid, approvals, structure validation, beam geometry, plan quality, and isocenter setup.

Figure 2. Prototype MOSAIQ reconciliation workflow integrated into the automated plan-check framework for end-to-end treatment workflow verification.

This work demonstrates the feasibility of a deterministic TPS-integrated automated plan-check framework for radiation therapy plan verification. By using standardized pass/fail criteria and transparent reporting, the system reduces reviewer variability and cognitive burden while providing rapid, reproducible QA directly within the TPS environment.

The automated checks are intended to augment clinical physicist review by improving clinical efficiency, consistency, and plan accuracy.

A deterministic automated plan-check system was developed within RayStation using the Python API. The framework performs more than 30 automated checks across multiple clinical domains and generates standardized HTML reports with transparent pass/fail reporting. The system demonstrated robust performance across more than 200 clinical treatment plans, reliably identifying predefined failure conditions with average runtimes under 3 seconds per plan. This framework improves consistency, reduces reviewer variability, and enhances patient safety through scalable, standardized plan-review automation.

Ford EC, et al. Strategies for effective physics plan and chart review in radiation therapy: Report of AAPM Task Group 275. Medical Physics. 2020;47(6):e236-e272.